

# CENTRAL INSTITUTE FOR COTTON RESEARCH

POST BAG NO 2, SHANKAR NAGAR POST OFFICE NAGPUR - 440 010

Tel. Phone: 07103-275536, Fax: 07103-275529, EPBAX: 07103 – 275637 / 38 / 39

{Location: Near Indian Oil Depot, Panjari, Wardha Road, Nagpur – 441 108}

F. No.(49)St/NFBSFARA/Autoclave/2011-12

Date: 10.01.2012

**Sub:** Quotation for the Purchase of **Autoclave** – reg.

Sir,

Sealed quotations are invited by the Director, Central Institute for Cotton Research for the purchase of following **Autoclave** on the terms and conditions prescribed below:

SN	Specification	Rate (In Rs.)	Qty.	Amount (In Rs.)
1.	<b>Autoclave Capacity: 175 liters</b> <b>(The detailed specification is enclosed)</b>	@ Rs.	ONE	Rs.

### Terms & Conditions:

1. The price should be inclusive of all taxes / VAT etc.
2. Sealed quotation along with the literature and pamphlets should reach this office on or before **31.01.2012**.
3. As the equipment/Instrument is urgently needed for our research work, the firms who will be able to make immediate supply F.O.R., Central Institute for Cotton Research, Panjari Farm, Wardha Road, Near Khapri, Nagpur should only quote the desired rates.
4. Envelops may please be super-scribed with the word "**Quotation for Autoclave**".
5. Instruments indigenous will be purchased from only those parties, who can ensure and undertake to provide the satisfactory after-sales-service within the country. The supplier should also ensure to offer a complete detailed package on back up support, prompt commissioning, after- sales-service including workshop backup, timely supply of spares, timely repair of equipments, proper training to Scientists/ Technicians on use.
6. Guarantee/Warranty period offered and after sales service available should be clearly mentioned.
7. The quotation should be quoted by solely local authorized stockiest Service distributor against this equipment, who has necessary service backup facility.
8. Authorized Certificate should be enclosed with the quotation for this equipment.
9. Those who are unable to provide the local service facility will not be entertained.
10. The make, quality, and specifications may be stated in detail/must be as per details given by this office. In case supply is not made within a fixed period the office reserves the right to cancel the supply order.
11. The payment will be made only after completing delivery according to the specifications, quality, and quantity given in the quotation/purchase order.

12. Other terms and conditions of supply if any may also be stated.
13. Equipment/Instrument will be accepted only after giving satisfactory demonstrations.
14. The Director, CICR, reserves the right to accept or reject one or all quotations without assigning any reason thereof.
15. The quotations, without the desired specification received after the due date are liable to be rejected.

**Administrative Officer**

**Specification for Autoclave**

1. Autoclave capacity: 175liters
2. Size: Height 80cm and width 60cm (size without stand) and width inside should be 55cm.
3. Double walled construction outer M.S./S.S. 304 and inner S.S. 304 with heavy Gauge & Argon walled.
4. Lid should be S.S. 304 heavy gauge dies pressed.
5. Equipped with pressure gauge, steam release valve and 2-safety valves
6. S.S. perforated basket
7. Working steam pressure should be 15PSI with 121°C
8. Silicon Rubber gasket joint less
9. Foot lifting arrangements to open lid
10. Pressure gauge on top with lid
11. Pressure cut-off switch
12. Low water level cut off system with alarm for safety, Water level indicator
13. Work on 230 Volts
14. Low power consumption
15. Warranty: 2 years