

CENTRAL INSTITUTE FOR COTTON RESEARCH
REGIONAL STATION, COIMBATORE 641003

Dr. N Gopalakrishnan
Project Co-ordinator & Head (Cotton
Improvement)

Fax: 0422 – 2454021
Phone: Off: 0422-2430045
Email: cicrcbe@gmail.com
Web: www.cicr.org.in

F.No.ST-I/1-1/Plan/08-09/

Dated:5.12.08

Sir,

Sub: Quotation for the purchase of PCR Thermal Cycler -reg.

Sealed quotations are invited by the Project Coordinator & Head, Central Institute for Cotton Research, Regional Station, Coimbatore-3 for the purchase of PCR Thermal Cycler as per the terms and conditions prescribed below:

Specification	Qty. Reqd.	Rate in Rs.	Amount (in Rs.)
PCR Thermal Cycler: Thermo block for 96 x 0.2 ml, and One 8 x 12 micro titer plate. Gradient Program should be for Annealing, Denaturation and Elongation temperature Gradient program up to 24 °C with LCD display Peltier element heating technology Programmable temperature control speed: Approx. 6 °C/s (heating) Approx. 4.5 °C/s (cooling) Initial ramp rate from room temperature should be very fast. Gradient temperature range of 30-99 °C. Ability to store methods on external card for safety and multiple user capability. Licensed and authorized system for PCR applications. External control panel for easy access and programming. Large VGA display with graphical display of temperature and time Future upgradeable to RtPCR Heated lid should have fluid cushion to minimize evaporation USB Port for data downloading 2 KVa online UPS (5 hours back-up) Data access system (SONY VAIO VGN-CR313H/B LAPTOP) - Quote the data downloading system separately	One No		

Terms & Conditions:

1. The price should be inclusive of all taxes/VAT
2. Sealed quotation along with the literature and pamphlets should reach this office on or before **20.12.08**
3. As the equipment/Instrument is urgently needed for our research work, the firms who will be able to make immediate supply to **Central Institute for Cotton Research, Regional Station, Coimbatore – 3** should quote.
4. Envelops may please be super scribed with the word **“Quotation for PCR Thermal Cycler”**
5. Instruments/Equipment will be purchased from only those parties, who can ensure and undertake to provide the satisfactory after-sales-service within the country. The supplier should also ensure to offer a complete detailed package on back up support, prompt commissioning, after-sales-service including workshop backup, timely supply of spares, timely repair of equipments, proper training to Scientists/Technicians on use.
6. Guarantee/Warranty period offered and after sales service available should be clearly mentioned.
7. The make, quality and specifications may be stated in detail/must be as per details given by this office.
8. **In case supply is not made within a fixed period the office reserves the right to cancel the order.**
9. **The payment will be made only after completing delivery according to the specifications, quality and quantity given in the quotation/purchase order with in 30 days.**
10. Other terms and conditions of supply if any may also be stated.
11. Materials will be accepted only after giving satisfactory demonstrations.
12. The Project Coordinator & Head, CICR reserves the right to accept or reject one or all quotation/s without assigning any reason thereof.
13. The quotations, without the above desired information/received after the due date are liable to be rejected.

Project Coordinator & Head